

## ***Interactive comment on “The Isotopx NGX and the ATONA Faraday Amplifiers” by Stephen E. Cox et al.***

**Daniel Wielandt (Referee)**

wielandt@bio.ku.dk

Received and published: 13 March 2020

General comments: This is a well-written article on an important novel amplifier technology called ATONA that provides an otherwise currently unavailable combination of low noise and high dynamic range for Faraday cup measurements of ion beams. The technology could significantly improve both current and future mass spectrometers, and is therefore of general interest to all mass spectrometry specialists. I however believe that its impact could be improved by including some additional information as mentioned below. Alternatively, the suggestions in general comments should be addressed in future publications.

The article focuses on comparing the ATONA to current 10E11, 10E12, 10E13 and

Printer-friendly version

Discussion paper



a hypothetical  $10E14$  ohm amplifier, or rather their idealized Johnson-Nyquist noise characteristics, for the purpose of multicollector noble gas measurements. ATONA outperforms ideal i.e. model  $10E13$  ohm amplifiers with respect to signal-noise ratio for 10 second integrations which is (most likely) an appropriate integration time for many measurements, and approaches an ideal (and currently commercially unavailable)  $10E14$  ohm amplifier for a 100 second integration which is most likely to long to properly sample and back-project a noble gas ion beam evolution to T0. The high dynamic range and low noise-fast response is definitely an improvement as compared to traditional amplifiers. This versatility means that amplifiers do not need not be physically or electronically switched among Faraday cups for different applications, which is an additional advantage that complements their low-noise characteristics. An ATONA could also be useful for single detector instruments that still have merit due to the high sensitivities afforded by the small volumes of such instruments.

Although the comparison with traditional amplifiers at low signal intensities is appropriate, the paper could benefit from a more stringent comparison with ion counters where the noise characteristics at low signal intensities are dominated by Poisson i.e. counting noise of the individual ion arrivals. This noise is inherent to counting atoms or ions and cannot be avoided. An interesting question is therefore under which beam intensity  $x$  time i.e. accumulated charge conditions the "baseline" noise in an ATONA becomes comparable to this inherent and unavoidable counting noise that will also be present and superimposed on zero-beam i.e. electronic baseline noise? This would seem to be an appropriate lower dynamic range where ion counters would (decisively?) outperform ATONA in terms of precision (but not necessarily accuracy). This number could presumably be calculated based on the 1-10-100 second zero-beam measurements that have already been carried out. It may also be possible to tease out that information from e.g. figure 7, but it is better that it is presented.

The paper would also benefit if the working principles of ATONA were more thoroughly discussed (without disclosing confidential information). The patent documents contain

Printer-friendly version

Discussion paper



a lot of public information that could be condensed into a description of the technology. I think the mass spectrometry community would be more likely to adopt the technology if they could understand it better, rather than using it as a "black box" technology where one might run into an unpredictable problem. As a naive non-engineer I personally would like to know how leakage current is reduced. Is there a maximum charge than can be accumulated before "discharging" if that is even the appropriate term? Are there hysteresis effects in the capacitor that make it particularly hard to drive out or sense low buildups of charge that might adversely affect linearity at low signal intensities? Can charge buildup in the Faraday-amplifier system start to deflect incoming ions, changing the peak shape thereby affecting e.g. pseudo-resolving peak-shoulder measurements. Does the "firmware" make decisions on sampling rate or readout parameters, switching between different regimes that depend on beam intensities?

Throughout: The term Johnson-Nyquist noise is used in line 114, but then subsequent usage is about Johnson noise. Should abbreviate it JN-noise at first usage, and then refer to it as such subsequently.

When discussing the performance using air and cocktail standards, it would be nice to have the approximate beam intensities tabulated in e.g. fA as that is the unit that is already reported for noise measurements.

Specific: First paragraph i.e. 8-22 could perhaps use a statement regarding engineering tradeoffs regarding multicollection versus volume/sensitivity, i.e. the increase in volume that tends to occur with multicollection and the related drop in sensitivity. This is one reason why single collector instruments still have a role. In fact, the versatility of the ATONA seems to make it very well suited for that role; this is only aided by its rapid response as discussed later.

Second paragraph, line 30. Mention of long settling time for high value resistors is relevant in case of dynamic measurements, but static multi-collection of noble gases all but removes the settling time issue since any single resistor only measures one

Printer-friendly version

Discussion paper



very slowly evolving beam. This should be mentioned in order to be fair to the current generation of high-ohm multi-collector equipped instruments.

Paragraph 6, Line 80. Could the authors perhaps make a back of the envelope error propagation calculation of how much of the air correction error on a blank subtraction on their instrument would arise from the  $^{36}\text{Ar}$  using a ion counter versus the ATONA? Or conversely the calculations suggested in the general comments regarding comparison of counting noise vs zero beam noise? This would be highly relevant for e.g. Ar or Ne dating of young samples where samples or fractions may be comparable in intensity to blanks.

Paragraph 7, Line 84. If possible, it would be nice if the patent were hyperlinked.

Paragraph 9 A formulation of Johnson-Nyquist noise with some appropriate reference and description would be useful for non-specialists.

Paragraph 11, Lines 130-140. This is a bit hard to read, and the reporting would benefit from a data-table showing the noise characteristics for 1, 10 and 100 second integrations with ATONA and 10E11-14 resistors. In such a way, one could focus on describing the noise "crossover" points for the various detector technologies that most readers would be searching for anyway as seen in figure 4.

Figure 6 (and figure A1) It is hard to identify the ranges, could the color code somehow be complemented by a change in marker style? It might also be a good idea to write the ranges as from 200% to 0.36% rather than between 200% and 0.36%.

Figure 7 We should expect a number of inflection points where all faraday mass spectrometer technologies gradually switch to follow a slope determined by counting noise ( $N^{0.5}$ ) rather than signal over "baseline" JN or kTC noise ( $N^1$ ). The linear error envelopes could give the erroneous impression that Faraday-based technologies can eventually outperform counting noise at high intensities, this should be avoided.

Table 1 The table should include the intensity of the smallest ion beam intensity i.e. the

Printer-friendly version

Discussion paper



$^{36}\text{Ar}$  intensity in fA. It would also be nice to have (and discuss) an MSWD to compare internal precision and external reproducibility for all the measurements. A calculated average for the different intensities would also be nice, and could be plotted to evaluate non-linearity.

Table A1 The table should include the intensity of the smallest ion beam intensity i.e. the  $^{38}\text{Ar}$  intensity in fA. It would also be nice to have (and discuss) an MSWD to compare internal precision and external reproducibility for all the measurements. A calculated average for the different intensities would also be nice, and could be plotted to evaluate non-linearity.

Data for 0.36% measurements seem improbably precise, are they missing a digit?

It would also be nice to discuss the presumably significant decrease in precision when going from 5.2% aliquots to 2.6% aliquots and lower. Is this a characteristic of ATONA, or is it due to error propagation effects from subtraction of blank  $^{38}\text{Ar} + \text{H}^{37}\text{Cl}$ ?

---

Interactive comment on Geochronology Discuss., <https://doi.org/10.5194/gchron-2020-1>, 2020.

Printer-friendly version

Discussion paper

