



Environmental Gamma Dose Rate Measurements using CZT Detectors

Sebastian Kreutzer^{1,*}, Loïc Martin², Didier Miallier³, and Norbert Mercier⁴

¹Institute of Geography, Heidelberg University, Im Neuenheimer Feld 348, 69120 Heidelberg, Germany

²Innsbruck Quaternary Research Group, University of Innsbruck, Innrain 52, 6020 Innsbruck, Austria

³Laboratoire de Physique de Clermont-Ferrand, Université Clermont Auvergne, Campus des Cézeaux, 24, avenue des Landais BP 80026, 63171 Aubière cedex, France

⁴Archéosciences Bordeaux, CNRS-Université Bordeaux Montaigne, Maison de l'Archéologie, Esplanade des Antilles, 33607 Pessac cedex, France

Correspondence: Sebastian Kreutzer (sebastian.kreutzer@uni-heidelberg.de)

Abstract. The accurate and precise determination of the environmental dose rate is pivotal in every trapped-charge dating study. The environmental gamma-dose rate component can be determined from radionuclide concentrations using conversion factors or directly measured in situ with passive or active detectors. In-field measurements with an active detector are usually inexpensive and straightforward to achieve with adequate equipment and calibration. However, despite the rather widespread use of portable NaI or LaBr₃ scintillator detectors, there is a lack of research on the performance and practicality of portable alternative detectors in dating studies, particularly in light of newer developments in the semi-conductor industry. Here, we present our experience with two small portable semi-conductor detectors housing Cadmium Zinc Telluride (CZT) crystals. Given their small volume and low power consumption, we argue they present attractive alternatives for gamma-dose rate measurements in dating studies. Despite high relative detection efficiency, their small volume may pose different challenges, resulting in impractical measurements in routine studies and, therefore need investigation. In our study, we simulated the particle interaction of the CZT crystal with *GEANT4* in different sediment matrices to quantify the energy threshold in the spectrum above which the count/energy-count rate correlates with the environmental gamma dose-rate irrespective of the origin of the gamma-photons. We compared these findings with experimentally derived cumulative spectra and dose-rate calibration curves constructed from reference sites in France and Germany, which yielded unrealistically low threshold values likely due to the limited variability of the investigated sites. We additionally report negligible equipment background and required minimal measurement time of only 20 min in typical environments. Cross-checking our calibration on a homogenous loess deposit near Heidelberg confirmed the setting and assumed performance through a nearly identical gamma-dose rate of $1107 \pm 65 \mu\text{Gy a}^{-1}$ (CZT) to $1105 \pm 11 \mu\text{Gy a}^{-1}$ (laboratory). The outcome of our study gives credit to our threshold definition. It validates the similarity of the two investigated probes, which may make it straightforward for other laboratories to implement the technique effortlessly. Finally, the implementation of CZT detectors has the potential to streamline fieldwork and enhance accuracy and precision of trapped-charge dating-based-chronologies.



1 Introduction

Asserting the effective environmental dose rate is indeed crucial for accurate and precise ages in luminescence and electron-spin-resonance (short: trapped-charge) dating studies. The dose rate plays a vital role in the age equation as it is a significant
25 factor in determining the amount of radiation absorbed over time.

Field procedures typically involve sampling sufficient bulk material around the sampling site. The material is then analysed to quantify the natural radionuclide concentrations (such as U, Th, and K concentrations) as major contributors to the environmental radiation. If collecting sufficient material is not feasible around the sampling site, e.g., in archaeological excavations, or drilled cores, the required material can be carefully separated from the to-be-dated material combined with in situ
30 measurements.

Strategies to ensure a good environmental dose rate estimation ideally interoperate both, laboratory and field, measurements. Standard analytical methods in the laboratory involve α -/ β -counting, γ -ray spectrometry (e.g., Aitken, 1985; Zöller and Pernicka, 1989; Hutton and Prescott, 1992; Preusser and Kasper, 2001; Godfrey-Smith et al., 2005; Mauz et al., 2021; Kolb et al., 2022) and element analytical methods such as inductively coupled plasma mass spectrometry (ICP-MS) (e.g., Preusser
35 and Kasper, 2001) or a combination of those methods. Activity or element concentrations are then converted for each type of radiation (α -, β -, γ -radiation) using dose-rate conversion factors (latest update: Cresswell et al., 2018).

Dose-rate (γ - and, more challenging, β -) components can be measured in the field at the sampling position using passive dosimeters (e.g., Hutton and Prescott, 1992; Kalchgruber et al., 2003; Kalchgruber and Wagner, 2006; Richter et al., 2010; Kreutzer et al., 2018b) stored over a couple of weeks to months or with active detectors (usually γ -ray probes) (e.g., Murray
40 et al., 1978; Mercier and Falguères, 2007; Guérin and Mercier, 2011; Arnold et al., 2012; Bu et al., 2021; Martin et al., 2024) enabling nearly instant dose rate estimates.

Regardless of the preferred method and type of detector, active or passive, in-field measurements appeal when sampling suggests a heterogeneous distribution of radionuclides or complex geometries (e.g., the close succession of very different sediment layers, gravels/rocks in the profile). The field dose-rates can later be compared to laboratory results based on the
45 radionuclide concentrations. Ideally, the obtained numbers agree within uncertainties, or the discrepancy gives further insight into the site's matrix composition. Active detectors can be paired with a portable luminescence reader (e.g., Sanderson and Murphy, 2010) to profile the stratigraphy and determine relative chronologies. A rule of thumb would approximate the γ -dose component as about 28% to 36% of the total dose rate (numbers derived from the ChronoLoess database by Bosq et al., 2023; alternatively, see estimates in Aitken, 1985). These numbers underpin the importance of the γ -dose rate contribution and its
50 significance in estimating accurate trapped-charge ages.

On the flip site, the usually short measurement durations, compared to the expected age of the sediment, has the disadvantage that long-term changes in the water content are not reflected. Contrary, passive dosimeters would register at least seasonal variations if stored over months in the field. Both (passive and active) do not register potential radioactive disequilibria. Furthermore, depending on the size of the detector probe a rather large hole with a depth of at least ca 30 cm is required for the
55 measurement. Such a hole is sometimes difficult to dig, not always possible (samples from a drilled core) or not favoured



given the setting (e.g., archaeological excavation). Here focussing on active detectors, additional everyday challenges involve equipment calibration and handling usually proprietary hardware such as cables or multi-channel analysers that are costly to repair or even unavailable after they have been phased out by the manufacturer. Last, the equipment can be bulky, especially for large detectors (up to 3×3 in for a portable NaI probe) and, given first-hand experience, the equipment is prone to preferred inspection during air travel.

In summary, while in-field measurements with active detectors do present certain challenges, their benefits are still considerable. They provide valuable, real-time data at a relatively low cost, significantly improving the accuracy of dating studies. As a result, their routine use seems advisable.

In the following, we will test two commercially available portable Cadmium Zinc Telluride (CZT) detectors for in situ γ -ray measurements. The detectors are small and highly portable, and we assume that they can pose an alternative to systems using larger NaI or LaBr₃ probes in trapped-charge dating applications. Next, we will begin outlining the technical specifications and advantages of the CZT systems. We will then detail the required calibration methods and explore the performance and dose-response characteristics of the detectors through simulations and measurements in different natural sites with well-known radionuclide concentrations. Finally, we will test the calibrated systems in a loess deposit near Heidelberg and discuss the results.

In this contribution, we focus exclusively on the “threshold” technique (Løvborg and Kirkegaard, 1974, further details below) for measuring environmental γ -dose rates ($\dot{D}\gamma$ in $\mu\text{Gy a}^{-1}$). Unlike the “window” (three windows each for U, Th, and K) method, which compares the area under a specific γ -peak in a sample with unknown composition to the area of a γ -peak in a sample with known radionuclide composition, the threshold technique integrates the entire spectrum above a set threshold. The threshold approach provides a direct measure of $\dot{D}\gamma$ in $\mu\text{Gy a}^{-1}$ rather than a radionuclide concentration for natural environments.

2 Material and methods

2.1 Brief background γ -detectors

Measuring γ -rays translates to observing the interaction of (γ -) photons with matter by quantifying the production of secondary charged particles. Suitable are scintillation detectors such as NaI(Tl) or LaBr₃. They collect light caused by the interaction of the γ -photons with the detector material or semiconductor-based detectors (e.g., high-purity Ge). Combined with a suitable electronics, they record the amount of produced secondary-hole pairs (e.g., Gilmore, 2008).

To measure γ -rays outside a laboratory, for instance, in trapped-charge dating studies, portable detectors that can be operated at room temperature are preferred. This usually favours scintillation detectors using NaI(Tl) or LaBr₃ with typical probes ranging from 1.5×1.5 in to 3×3 in over HPGe semiconductor-based detectors that require operation at liquid nitrogen temperature due to small band-gap of the crystal. Cadmium Zinc Telluride (CZT and CdZnTe) were proposed as promising alternatives with better γ -ray absorption performance and operational at room temperature. However, the production process



is more challenging (e.g., Gilmore, 2008) and such detectors were not an option considered in the context of trapped-charge geochronology; yet.

90 Since the 1990s, the development of CZT semiconductor detectors progressed considerably in their applicability as γ -ray detectors (for reviews, see Scheiber and Chambron, 1992; Verger et al., 1997; Limousin, 2003; Alam et al., 2021). They offer a small volume and operate at ambient temperature by collecting charges created by the interaction of ionising radiation with a high relative efficiency for photoelectric interaction (atomic numbers Cd: 48, Te: 52; density crystal ca 5.8 g cm^{-3}) (Limousin, 2003; Alam et al., 2021). Although this is less important in our case, they provide an energy resolution comparable to or better
95 than LaBr_3 and considerably higher than NaI(Tl) (Alexiev et al., 2008) probes. Also considering the small volume available for detection, their absolute efficiency remains lower than that of larger detectors. This feature, combined with a low energy consumption, renders this detector type particularly appealing for our application.

2.2 Equipment

For our experiments, we used two systems from Kromek (<https://www.kromek.com/>; last access: 2024-08-17) with CZT detectors. (1) RayMon10[®] (henceforth: RayMon GR1) and (2) GR1+[®] (henceforth: GR1)¹ (Fig. 1). Both systems include a similar
100 $10 \times 10 \times 10 \text{ mm}^3$ GR1 CZT detector connected to a 4096 energy/channels analyser. The detection ranges from 30 keV to 3 MeV with an energy resolution of around 2.5% FWHM at 662 keV. The RayMon GR1 was delivered with a handheld touch-screen device running Microsoft Windows 10[®] and comes housed. The probe communicates with the handheld device via a Universal Serial Bus (USB) Type A connector. The battery lasts around eight to ten hours, depending on the display bright-
105 ness setting. Although much smaller in housing size (GR1: $25 \text{ mm} \times 25 \text{ mm} \times 63 \text{ mm}$, 60 g; compared to RayMon GR1: $42 \text{ mm} \times 35 \text{ mm} \times 180 \text{ mm}$, 176 g), the GR1 contains a similar CZT detector. It has a Mini-A USB port that can be attached to any standard computer given a suitable cable and operated using the software K-Spect[®] that can be downloaded free of charge from the manufacturer. The GR1 consumes only 250 mW and is hence operational as long as the battery of the connected computer lasts. For more information, we refer to the manufacturer's website.

110 Because the Mini-USB port of the GR1[®] seemed fragile, and we were not sure about the sealing of the housing against moisture, we designed a 3D-printed, rubber-sealed strain relief mount (Fig. 1) and attached it to the detector housing. The strain relief enables safe retrieval of the detector, and the simple plastic bag wrapping keeps dirt and moisture away during field operations. The adapter was designed by the Scientific Workshop Service of Heidelberg University and we share the print-ready files under CC BY-NC 4.0 licence conditions on Zenodo as a supplement to this article.

115 2.3 Calibration methods

We aim to use the detectors in routine dating applications to determine \dot{D}_γ in $\mu\text{Gy a}^{-1}$. This requires three separate experiments in given order to set up each device: (1) Channel/energy calibration, (2) energy threshold definition, and (3) a calibration curve modelling counts against the environmental dose rate.

¹The plus indicates a slightly higher energy resolution compared to the "non-plus" GR1 version.



Figure 1. Kromek detectors used for our measurements (shown is the probe without the handheld tablet PC for the RayMon10[®]). Both probes house a similar CZT detector. We wrapped the GR1[®] in a standard plastic bag and attached a home made strain relief to the GR1[®] to enable easier operation and retraction of the detector in the field.

The dead time was insignificant during all the experiments presented in this manuscript. The dead time is the difference
120 between real-time and live time, which equals the time when the detector did not register new counts. The longest dead time of both detectors was 3.6 s with the highest relative dead time amounting to only 0.1 %.

2.3.1 Channel/energy calibration

The channel/energy calibration (Sec. 3.1) assigns energy values in keV to the, in our case, 4096 channels. The calibration
125 makes it easier to interpret the γ -ray spectrum, enables a comparison of spectra, and accounts for shifts in the spectrum that may occur due to, for instance, changed environmental conditions.

Both detectors used here were delivered with a test and inspection sheet documenting measurements against ²⁴¹Am (γ -line at 59.5 keV) and ¹³⁷Cs (γ -line at 662 keV). The results are nearly identical for both detectors with an offset of ca 2 channels/keV between the GR1 and the RayMon GR1.

For the channel/energy calibrations, where only the peak position matters, we used two γ -standards available in Heidelberg
130 closely arranged around the detector for two measurements over 3600 s. One radiator is a home-made uranium standard (U concentration: 1.02%) and the other an Amersham EB 165 mixed radionuclide standard with ²⁴¹Am and ¹³⁷Cs. The Amersham standard also contains other shorter-lived radionuclides; however, given the age of the standard (>30 years), we do not expect to observe significant counts above background within the chosen measurement time.



2.3.2 Energy threshold determination

135 The energy-threshold definition (Sec. 3.5) determines the threshold in the spectrum above which \dot{D}_γ is seemingly independent of the origin of the absorbed γ -photons (see Løvborg and Kirkegaard, 1974, for details). In other words, the integrated spectrum above the threshold is used to derive \dot{D}_γ . Guérin and Mercier (2011) distinguished two different thresholds techniques for *integrating* the spectrum. The “count” and the “energy” threshold (integration technique). The *count* threshold adds all counts above a certain threshold (η) whereas the *energy* threshold integrates the deposited energy above η . Assuming that S_i is the
140 signal registered either as absolute counts per channel or count rate per channel (s^{-1}) in the i^{th} channel of the spectrum, E_i in keV the energy associated with a certain channel. The relationship between the environmental γ -dose rate and integrated value above the threshold for an energy/channel calibrated spectrum becomes, in case of the counting threshold technique:

$$\dot{D}_\gamma \sim \sum_{i:\geq\eta}^N S_i \quad (1)$$

and it reads

$$145 \quad \dot{D}_\gamma \sim \sum_{i:\geq\eta}^N S_i \times E_i \quad (2)$$

for the energy threshold integration technique. Guérin and Mercier (2011) found η slightly lower for the latter technique, resulting in a larger proportion of the spectrum usable, which lowers the statistical uncertainty. Although related, the two threshold integration techniques must be distinguished from quantifying η , i.e. finding the energy above which \dot{D}_γ is a function of the integrated counts (Løvborg and Kirkegaard, 1974), regardless of the integration technique. To determine η , one can
150 perform energy-matter interaction simulations (e.g., Guérin and Mercier, 2011) or measure the γ -ray spectra of “pure” emitters of known U, Th, K concentrations (Mercier and Falguères, 2007; Rhodes and Schwenninger, 2007; Duval and Arnold, 2013).

For our contribution, we modelled the threshold (henceforth: η_{sim} in keV) with *GEANT4* (Agostinelli et al., 2003) using three different sediment matrices: (1) a pure SiO_2 matrix, (2) a brick-like matrix (SiO_2 : 66%, Al_2O_3 : 18%, Fe_2O_3 : 6%), and (3) a calcite rich sediment (CaCO_3 : 60%, SiO_2 : 40%). We set the matrix densities to 1.8 g cm^{-3} , and added no water (dry
155 matrix). The simulation geometry represents the RayMon10[®] probe according to the available documentation (Fig.2). The probe was placed at the centre in a 160 cm distance next to the cube containing the sediment matrix, ensuring an infinite γ -radiation matrix around it. The γ -ray emissions were simulated from each matrix using *GEANT4* electromagnetic physics from the G4PenelopePhysicsphysics constructor (Ivanchenko et al., 2011), which is based on the 2008 version of the *PENELOPE* Monte Carlo code for low energy particles (Baró et al., 1995). This *GEANT4* physics was successfully already tested for
160 simulating γ -photons from natural radionuclides in Guérin and Mercier (2011), Guérin and Mercier (2012), Martin et al. (2015).

The γ -spectra of ^{40}K , the U-series (in secular equilibrium) and the Th-series (in secular equilibrium) were built from the data of ENSDF database as of June 2014 (<http://www.nndc.bnl.gov/ensarchivals/>; last access: 2024-09-15) to independently simulate 1 Gy of γ -dose in the matrices. In the simulation, we recorded the energy of each γ -interaction with the CZT crystal,




Figure 2. Geometry of the simulated CZT detector – image from *GEANT4* visualization. Legend: (1) Sediment matrix; (2) rubber coating; (3) aluminum shielding; (4) air; (5) plastic support; (6) CZT crystal; (7) electronic components. Information kindly provided by the Kromek Group plc.

165 and the spectra of counts per energy channel were built for each matrix and each γ -emission spectra. These “measured” spectra
obtained by simulation were then used to create the curve of counts/deposited energy above the energy thresholds ranging
from 0 keV to 1000 keV. We then compared the standard deviation between the ^{40}K , U-series and Th-series curves of counts
above the threshold to quantify the optimal threshold for which the number of counts/energy above is proportional to the dose
rate and mostly independent of the natural radionuclide composition. We did not consider the sensitivity variation of the CZT
170 detector with energy and possible dead times. However, we assumed these phenomena have a low impact on determining the
count/energy threshold.

We compare these findings with measurements at five calibration sites (Fig. 3) to derive η_{exp} . Four sites are located in France,
three in the vicinity of Clermont-Ferrand (France) (Miallier et al., 2009), and the home-made brick block in the cellar of the
Archéosciences Bordeaux laboratory (Richter et al., 2010). Another site is a granite block (FLOSSI) located at the Max Planck
175 Institute for Nuclear Physics (Heidelberg, Germany). The granite block was donated by the Granitwerke Leonhard Jakob KG to
the Forschungsstelle Archäometrie (Günther A. Wagner) in 1991 for the purpose of having reference site for calibrating γ -ray
spectrometers. The radioelement concentration of the block was analysed with neutron activation analysis, atomic absorption
and high-resolution γ -ray spectrometry as part of the work by Rieser (1991). Although the information from this analysis was
later used by others (e.g., Hossain et al., 2002; Kalchgruber, 2002) the values were never formally published. We therefore
180 added the CSV file with the values from Rieser (1991) to our Zenodo dataset (Kreutzer et al., 2024).

In general, the investigated sites (Fig. 3A-E) have a well-known radionuclide composition from which \dot{D}_γ can be calculated
to construct γ -dose rate calibration curves using the two threshold integration techniques to re-evaluate η as the value where
the mean square of residuals from the model reaches the lowest value. The underlying assumption of this approach is that if the
threshold was set correctly, the regression line should exhibit the best fit as a non-ideal threshold should increase the residuals
185 due to a poor fit. Insufficient model adaptation is caused by poor counting statistics (threshold too large) or in situations where
the prerequisite of the technique that \dot{D}_γ is independent of the origin of the γ -photons is not fulfilled (threshold too low). We




Figure 3. Photos of all natural sites measured in this study. A-E are calibration sites with known radionuclide composition, (F) is a loess deposit at the Weiße-Hohl near Heidelberg (Germany) used to cross-check the equipment calibrations. Details to A-D can be found in Miallier et al. (2009) and Richter et al. (2010), details to E and F are provided in the maintext. The red circles mark the measurement positions (holes) for the probe. The sites BDX (D) and FLOSSI (E) are located in areas with restricted access. (D) in the basement of the Archéosciences Bordeaux laboratory at the Université Bordeaux Montaigne in Pessac (France) and (E) at the Max Planck Institute for Nuclear Physics in Heidelberg (Germany).

will compare those values with the experimental approach of deriving the threshold, except that we do not have access to sites with pure radionuclide concentrations but will use measurements from sites Fig. 3A-E instead.

2.3.3 Dose-rate calibration curves

190 The dose-rate calibration curve (Sec. 3.6) correlates the integrated (count and energy integration technique) signal with \dot{D}_γ from the reference sites (Table 1), i.e. the response of the detector to natural γ -radiation. If established, it allows us to derive an accurate estimate of \dot{D}_γ from a natural site with unknown radionuclide composition. As pointed out by Guérin and Mercier (2011), the water content will not affect the counting rate significantly, and the established value should be applicable to sites usually probed in trapped-charge dating applications.

195 The \dot{D}_γ values in Table 1 differ from the values reported in Miallier et al. (2009) after we recalculated them using the conversion factors compiled by Cresswell et al. (2018). Values recalculated for other conversion factors can be found in the dataset `clermont_2024` contained in the R 'gamma' package (> v1.1.0).



Table 1. Known γ -dose rates from reference sites. The dataset listed here, except for FLOSSI, ships with the R package ‘gamma’. In the original dataset, ‘BDX’ is termed ‘BRIQUE’, which is the brick block in the Archéosciences Bordeaux laboratory; for clarity, we relabeled it to BDX for our analysis. The values for FLOSSI represent the central values and standard error (Galbraith and Roberts, 2012) of all respective analyses given in Rieser (1991). We list the results calculated with the conversion factors by Cresswell et al. (2018). Please note that data in Miallier et al. (2009) are given as total dose rate, including the cosmic-dose rate contribution. Here we can neglect the cosmic dose rate contribution as we cut the spectra at 2800 keV.

SITE	NATURE	U	σ_U	Th	σ_{Th}	K	σ_K	\dot{D}_γ	$\sigma_{\dot{D}_\gamma}$
BDX	ceramic	4.1	0.1	13.7	0.4	3.5	0.1	1997.1	37.7
C341	trachybasalt	1.8	0.0	6.4	0.4	1.4	0.0	855.2	21.8
C347	granite	2.8	0.1	4.7	0.1	3.5	0.1	1425.5	27.4
FLOSSI	granite	19.2	0.6	13.4	0.4	4.1	0.1	3797.4	92.5
PEP	granite	6.0	0.2	19.0	2.0	3.8	0.2	2554.3	112.7

Note:

U, Th concentrations in $\mu\text{g g}^{-1}$, K in % | dose rates in $\mu\text{Gy a}^{-1}$.

Please note that for establishing the calibration curves we assumed “infinite matrix” conditions that enabled us to convert the radio-nuclide concentrations into dose rates (e.g., Guérin et al., 2012, for a critical review of this concept).

200 2.4 Radionuclide determination cross-check

To validate our calibration and post-processing procedure, we recorded natural γ -spectra at the Weiße Hohl (WH2024). The site is a gully of anthropogenic origin that cut into the famous last-glacial aeolian deposits near Nussloch (Germany) (e.g., Antoine et al., 2001). Today, the gully is part of a hiking trail in the area and hence easily accessible. The Nussloch loess deposits are well-investigated through numerous studies, and the expected \dot{D}_γ at Weiße Hohl was about 1 Gy ka^{-1} (Rieser, 205 1991). What made the measurements at this particular site interesting was that loess is typically subject to past climate and chronology studies using trapped-charge dating methods and reflect an often encountered use case.

We recorded two spectra over 20 min with both detectors in a 32 cm deep hole and sampled about 120 g of material for subsequent radionuclide and gravimetric water content quantification. We further extracted two subsamples for radionuclide concentration analyses in Heidelberg and Bordeaux. In Heidelberg, we employed a μDose (Tudyka et al., 2018; Kolb et al., 210 2022) and a $\mu\text{Dose+}$ (Tudyka et al., 2024) system on the same 3 g subsample. The sample was measured more than two days each. On another 83.3 g, we performed high-resolution γ -ray spectrometry measurements in Bordeaux (Guibert and Schwoerer, 1991). To compare the \dot{D}_γ calculated from the radionuclide concentrations using the conversion factors by Cresswell et al. (2018). We corrected the \dot{D}_γ measured with GR1 and RayMon GR1 for the field water content (Aitken, 1985). As for the calibration measurements, we assumed “infinite matrix” conditions and approximated a 4π geometry.



215 2.5 Data and data processing

We used *GEANT4* (Agostinelli et al., 2003) for the threshold modelling and processed our data with R (R Core Team, 2024) and the packages 'gamma' (Lebrun et al., 2020; Frerebeau et al., 2024) and 'ggplot2' (Wickham, 2016). The two investigated Kromek measurement systems provide export functionality for various data formats. We opted for the ASCII format `.spe` and added support in the function `gamma::read()` to the package 'gamma' (>v1.1.0) for this study (Frerebeau et al.,
220 2024). Except additions detailed below, our workflow uses the analysis functions of the 'gamma' package and follows the suggestions by Lebrun et al. (2020) and the tutorials that come with the 'gamma' (>v1.1.0) R package. This also includes the steps to determine the dose-rate response curve. For clarity, it should be mentioned that the 'gamma' package internally uses the function `IsoplotR::york()` (Vermeesch, 2018) to implement a regression analysis with correlated errors of xy-values that have individual uncertainties (York et al., 2004).

225 To ensure that the figures have colour-blind-friendly colours, we used the R package 'khroma' (Frerebeau, 2024) and the manuscript was prepared with 'rticles' (Allaire et al., 2024). A shortened version of the R code used for all the calculations, data, and calibration output are available on Zenodo (Kreutzer et al., 2024) under CC BY 4.0 licence conditions in accordance with common data-sharing guidelines.

3 Results

230 3.1 Energy calibration

Figure 4 shows the spectrum plot of GR1 measured over 1 h. We placed our γ -standards with known composition in front of the detector. The dashed lines marked the γ -lines used for the channel/energy calibration. The inset draws the energy/channel-calibration curve applied subsequently to all analysed spectra. We did not apply low-level discrimination and recorded raw count values for the measurements, i.e. the count rate was calculated in the post-processing. We have chosen the measurement
235 time to achieve a good counting statistic.

Given the nuclide composition, we expected to see typical γ -lines present in the ^{238}U decay chain on top of ^{241}Am and ^{137}Cs . The manufacturer also used the latter two nuclides before delivery to test the CZT detectors' performance and hence they provide a good reference for a cross-check. We manually identified eight γ -lines in our spectrum and assigned the results to the imported spectra with the `gamma::energy_calibrate()` function. To ease the peak identification, we started with
240 the ^{241}Am and ^{137}Cs γ -lines for which we have channel-to-energy references determined by the manufacturer. For instance, the manufacturer specifies to find the ^{241}Am peak @ 59.5 keV in channel number 80 ($\pm 10\%$) and the ^{137}Cs peak at 662 keV at 880 ($\pm 1\%$). Our calibration confirmed those values with channel number 82 for ^{241}Am 59.5 keV and channel number 877 for ^{137}Cs 662 keV.

The same calibration was performed with the RayMon GR1 detector but with its measurement time reduced to 900 s as a
245 cross-check. According to test data shared by the manufacturer on request (personal communication via e-mail, 2024-09-27), peak areas do not differ by more than 5% to 10% if the equipment is operated within the specified range (0-40°). In our case,




Figure 4. Energy calibration results for detector GR1. The main plot shows the raw spectrum with known γ -lines marked with dashed lines. The inset displays the calibration curve applied to all subsequently shown spectra. Peak positions were found to be similar for GR1 and RayMon GR1.

we found that the peak positions of RayMon GR1 were virtually identical to GR1 also under different temperature conditions (ambient temperatures at Clermont Ferrand: ca 28°, at FLOSSI: ca 18°; data not shown). Hence, for simplicity, we applied the GR1 channel/energy calibration to all measured spectra, and all spectra shown subsequently are energy/channel-calibrated.

250 3.2 Background measurements

To investigate the detector's counting background, we placed the GR1 for ca 5 h (18060 s) in a lead housing inside a low-level background environment at the PRISNA facility (Plateforme Régionale Interdisciplinaire de Spectroscopie Nucléaire en Aquitaine) near Bordeaux. The facility is a scientific platform used for low-level γ -ray spectroscopy experiments. Figure 5 illustrates that the system background is insignificant compared to typical environmental situations. The average count rate of the sum spectrum amounts to only 0.1 s⁻¹. Given the similarity of both detectors, we applied the same background subtraction to results from both detectors.

3.3 Energy-calibrated raw spectra

Figure 6 displays all energy calibrated γ -ray spectra measured at the sites at Clermont-Ferrand, Bordeaux, Heidelberg and PRISNA. We show count rates instead of absolute count values to account for the different measurement time. The shortest




Figure 5. Background measurements with GR1 in the lead castle for more than 5 h. The system background is negligible compared to typical measurements in the field.

260 live time was 1200 s (PEP) and the longest 18059 s (PRI). All spectra in Fig. 6(A) and (B) are scaled and colour-coded similarly for better comparison. Site PRI (the background measurement) was only measured with the GR1.

Visible in both spectra is a dominant Compton continuum rather than distinguishable photo peaks. This observation is not surprising given the short measurement time, the low abundance of the radionuclides (e.g., Miallier et al., 2009, for the Clermont-Ferrand sites) and, of course, the relatively low absolute efficiency for the small CZT crystal. It is reassuring that all
265 comparable raw spectra appear very similar in intensity, position and shape, except for the C341 spectra.

The spectra recorded in site C341 (a basaltic rock) appear to show only half of the counts measured with detector GR1 compared to the RayMon GR1 detector. This discrepancy is because GR1 was controlled via an external mobile computer that went unexpectedly into sleep mode. After reactivating the computer, the software seemed to have continued counting. However, post-processing revealed that it had stopped registering γ -photons. In other words, the difference between the two
270 readings (GR1 vs RayMon GR1) for C341 is a technical error, and hence, we discarded the spectrum C341 measured with GR1 for subsequent analysis. This error can be avoided easily but we kept it in the manuscript to share our experience.

3.4 Minimum required measurement time

When we performed our measurements at the reference sites, we still needed more practical experience with the two detectors. Therefore, we opted for measurement times longer than the typical setting for the LaBr₃ probes (ca 10 min). Unfortunately,
275 hour-long measurements for one sampling spot are often impracticable, considerably reducing the practicability of in-field measurements.

To assess the reasonably required measurement time for recordings, defined as a stable count rate within uncertainties in the field, we placed the RayMon GR1 detector in the brick block at Bordeaux (site: BDX) and started measurements for 60 s, 300 s, 900 s, and 3600 s (Fig. 7). Given the similarity of both detectors, we assume that this experiment will also be valid for




Figure 6. Channel/energy calibrated spectra as recorded in the reference sites as count rate against energy. (A) Spectra measured with detector GR1, (B) spectra measured with detector RayMon GR1. The spectra for both detectors are virtually identical in terms of peak position and count rates. The count rate for spectra C341 is significantly lower in (A) compared to (B). This is due to a software error (see maintext) and therefore this spectrum was discarded for later analysis. ‘PRI’ refers to the background spectrum recorded in the lead castle. We limited the x-axis to energy range later used for the integration: 30 – 2800 keV.




Figure 7. Sum of counts normalised to the measurement time recorded in the brick block at Archéosciences Bordeaux. After 20 min the average count rate does not change anymore within uncertainties. The plot scales depending on the settings of η (the energy threshold) and it was arbitrarily set to 200 keV for this graph.

280 the GR1. In the post-processing we integrated all spectrum counts for the experiment using the integration settings given below and normalised them to the measurement duration. The count rate is a little bit erratic over the first 500 s before stabilising after 20 min of measurement time. Additional measurement time increases the count rate only slightly. We therefore conclude that 20 min suffice in typical environments to determine a stable signal. This time corresponds to a total number of ca 4500 counts (4370counts).

285 3.5 Threshold definition

In Sec.2.3, we outlined the concept for defining the optimal energy threshold (η) above which the count rate correlates with the absorbed dose, regardless of the nature of the emitter and the matrix composition. The threshold is, in essence, a function of particle interaction with the (CZT) detector. Løvborg and Kirkegaard (1974) estimated the energy threshold for their setup (3×3 in NaI detector) at 500 keV, Murray et al. (1978) settled on 450 keV for their 2 in diameter NaI(Tl) probe. Mercier and Falguères (2007) calculated a threshold of 320 keV for their 1.5×1.5 in NaI(Tl) probe, a value later largely confirmed by simulations by Guérin and Mercier (2011) (their threshold value: 296 keV). Also, Duval and Arnold (2013) reported comparable values for NaI and LaBr₃ detectors of the same sizes (LaBr₃(Ce): 358 keV, NaI(Tl): 322 keV). As a rule of thumb, the larger the detector (higher absolute efficiency), the further the threshold shifts to higher energies. Since the volume of our CZT detector is small, we would locate the threshold in the low-energy part of the spectrum not higher than the values to be found
295 in the literature.



3.5.1 GEANT4 simulations

Figure 8A-B exhibits the simulations results for the three different matrices. We show the relative standard deviation between the number of counts above the energy threshold recorded during the simulations of 1 Gy generated with spectra of the radionuclides of the U-series, Th-series and the ^{40}K . This standard deviation is minimized when the number of counts/energy above the threshold is less dependent on the radionuclide of a chain of origin of the γ -photons, i.e. when the number of counts above the threshold is proportional to the dose absorbed by the detector, despite of the origin of the γ -rays of natural origin. The minimum standard deviation, obtained between 192.5 keV and 242.5 keV for the count integration technique and 97.5 keV and 222.5 keV for the energy integration technique, correspond to the curves in Figs. 8A-B falling below 10 % of the relative standard deviation (horizontal dashed line in Figs. 8A-B). This represents the optimal energy range for setting the energy threshold for the detector η_{sim} according to our simulations for the two integration techniques respectively.

3.5.2 Field measurements

3.5.3 Classical method

The “classical” method to determine η_{exp} experimentally, are measurements in environments with different and ideally pure radionuclide compositions. Here we tried to determine the energy threshold using the calibrations sites at hand for the counts and the energy integration technique (Figs. 8C-D). The threshold is defined as the smallest relative standard deviation of all spectra normalised to the respective environmental γ -dose rate of the sites. In Figs. 8C-D we only show the results of the detector RayMon GR1.

For the counts integration technique, we obtained η_{exp} at 98.8 keV and at 59.4 keV for the energy integration techniques. Both values are considerably smaller than the results from our simulation. We will later show that the simulated η is likely more accurate than the experimentally derived one. We attribute the difference to the similarity of the measured sites. Although \dot{D}_γ varies for all sites, the U/Th ratio is rather similar, likely leading to an unrealistically low value of η .

3.5.4 Calibration curve fitting

As an alternative to the simulation and experimental quantification of, η_{exp} , we experimented with a different approach. Therefore we calculated the γ -dose rate response curve for different energy windows using `gamma::dose_fit()`. Amongst other values, the function returns the mean square of the residuals (MSWD), which we can use to approximate the quality of fit of our regression model. Values lower or higher than 1 indicate a poor model adaptation. We defined the moving lower energy limit as E_i ($i := \{30, \dots, 1000\}$) and E_{max} was set to 2800 keV to avoid counts from cosmic-rays. Figures 8E-F show the outcome of this calculation for both detectors (blue: GR1, red: RayMon GR1). Although the curve of the mean residuals differ, the divergence of the determined thresholds are small and we believe that this deviation is caused by the discarded data point C341 for GR1, which is the lower point in the calibration curve.



For the count *calculation* technique (not shown in Fig 8E), the minimum in the search window between 30 keV and 350 keV was found at 91 keV. For the energy counting *calculation* technique, we located the value at 71 keV. Also these values are smaller than their simulated equivalents and the approach cannot compensate for the lack of differences between the measured sites. We therefore decided to continue with the simulated energy threshold values (i.e. $\eta := \eta_{sim}$)

330 3.6 Dose-rate calibration

With the threshold η derived from η_{sim} , we can obtain our dose-rate model again with the function `gamma::dose_fit()`, but this time for a fixed count/energy threshold at 99 keV and 59 keV, respectively. The results are shown in Fig.9 for the detector GR1 (Fig.9A) and RayMon GR1 (Fig.9B). Visual inspection confirms a good fit of the model to the data. However, the calibration curves differ slightly between the two detectors, which is likely due to the lower number of available data points.

335 The fitting parameters of both regression lines overlap with uncertainties (standard error as sum of weighted deviation from the fit; see York et al. (2004)). Acknowledging minimal variations between the CZY crystals and differences in housing and electronic, a perfect match is, however, not expected.

The package 'gamma' automatically fits the data for the energy threshold *calculation* technique and the counting threshold *calculation* technique for a given η . In Fig.9 we have shown only the latter. However, both values are accessible and saved
340 in the file `CAL_heiLUM_V0.rda` we made accessible at Zenodo (Kreutzer et al., 2024). As a reminder, we converted the radionuclide concentrations from the reference sites to dose rates using conversion factors compiled by Cresswell et al. (2018). These values influence the slope and intercept of the calibration curves. Because it may be desirable to apply additional calibrations based on other available conversion factors we repeated the calibration using conversion factors from Adamiec and Aitken (1998), Guérin et al. (2011), and Liritzis et al. (2013) (see data on Zenodo: (Kreutzer et al., 2024)).

345 3.7 Cross-check against natural site

The measurements at the Weiße-Hohl confirm once more that the two detectors exhibit very similar characteristics in terms of count rate efficiency (Fig.10). Differences seem stochastic without visible systematic diversion over the measurement duration of 20 min. To estimate the uncertainties, we implemented a new routine in the 'gamma' R package (argument: `dose_predict(..., use_MC = TRUE)`) that uses a Monte Carlo simulation approach, re-sampling from distributions
350 for slope, intercept, and the signal to predict the dose rate on the regression line. We found that with this method, the uncertainties increase of the already implemented analytical approach, however, it should reflect the true uncertainties more realistically.

The water content from the sample site (sample code: WH2024) was estimated at 2.1 % in the laboratory and this value was used to correct \dot{D}_γ . Table2 summarises the derived dose rate results for the two threshold integration techniques. $\dot{D}_{\gamma-final}$ is the arithmetic average of the values of these two techniques. The results for GR1 and RayMon GR1 agree within 2σ
355 uncertainties. This observation is likely caused by the calibration of GR1 sitting on fewer data points. The comparison of \dot{D}_γ against values derived from the radionuclide concentrations on WH2024 demonstrate a good agreement with field measurement uncertainties. If we compare the CZT results (GR1 and RayMon GR1) with the laboratory derived \dot{D}_γ , both summarised as




Figure 8. (A)-(B) *GEANT4* simulation results. (C)-(D) cumulative γ -ray spectra for all natural calibration sites normalised to the respective environmental γ -dose rate. Data only shown for RayMon GR1. (E)-(F) A variant of the experimental data as mean square of residuals (MSWD) of the dose-rate model fitting against the chosen minimum energy threshold. The values at 1 would indicate the best fit. The calculated energy threshold (η) is indicated in each of the plots as dotted line. Please keep in mind that while the three sets of graphs aim at showing the energy threshold using different methods, they neither represent the same data nor the same fitting method. For more details see main text.




Figure 9. Dose-rate calibration curves for detector GR1 (A) and RayMon GR1 (B). Uncertainties are given as standard errors (for details see York et al., 2004). Shown is the known γ -dose rate from the reference sites against the integrated energy signal between the threshold η (in keV) and 2800 keV. Note: The graphs give the impression that the uncertainties for the sites measured in France increase proportional to the size of the known γ -dose rates. This is a coincidence for our data subset and the effect not a real (cf. Miallier et al. (2009)).




Figure 10. Gamma spectra recorded at Weiße Hohl (Germany) with the two detectors. The inset shows that the absolute count rate difference between the two detectors is randomly distributed.



Table 2. Dry γ -dose rate results for the sample WH2024 obtained with different methods. Uncertainties on the final dose rate are quoted in 2σ approximating the 95 % confidence interval. The dose rates for the CZT estimates include an systematic error contribution of 1% from the energy calibration and was corrected for the in situ water content. Beyond digits listed here, we calculated with the full precision as returned by the measurement devices. The CZT measurements are in situ measurements, the μ Dose devices sampled 3 g of material each and the γ -ray spectrometer measurement used 88.3 g. REF_Central is the central value and its uncertainty from the laboratory derived \dot{D}_γ values.

DETECTOR	U	σ_U	Th	σ_{Th}	K	σ_K	$\dot{D}_{\gamma-Ni}$	$\sigma_{\dot{D}_{\gamma-Ni}}$	$\dot{D}_{\gamma-NiEi}$	$\sigma_{\dot{D}_{\gamma-NiEi}}$	$\dot{D}_{\gamma-final}$	$\sigma_{\dot{D}_{\gamma-final}}$
GR1	NA	NA	NA	NA	NA	NA	1095.35	185.98	1063.28	170.82	1105.15	258.33
RayMon GR1	NA	NA	NA	NA	NA	NA	1084.00	109.79	1079.84	92.26	1107.82	146.80
μ Dose+ (05)	2.58	0.37	10.91	0.89	1.18	0.03	NA	NA	NA	NA	1115.65	36.27
μ Dose (25)	3.14	0.36	9.72	0.84	1.23	0.06	NA	NA	NA	NA	1132.54	34.25
HPGe Prisma	3.10	0.02	8.70	0.07	1.20	0.01	NA	NA	NA	NA	1069.03	2.92
REF_Central	3.09	0.02	9.44	0.53	1.20	0.01	NA	NA	NA	NA	1105.30	21.25

Note:

U, Th concentrations in $\mu\text{g g}^{-1}$, K in % | dose rates in $\mu\text{Gy a}^{-1}$. | Prisma is a γ -ray spectrometer of the Archéosciences Bordeaux

central values (e.g., Galbraith and Roberts, 2012) we obtained $1107.2 \pm 65.1 \mu\text{Gy a}^{-1}$ (CZT) and $1105.3 \pm 10.8 \mu\text{Gy a}^{-1}$ (laboratory).

360 The results indicate a good homogeneity of the site reflected in the agreement between the field and the sampling dose rate. To get a better feeling for the sensitivity of \dot{D}_γ as a function of η for our detectors, we can calculate \dot{D}_γ of WH2024 for different values η . For this experiment on the energy calibrated spectra, we have to repeat the dose-rate calibration curve fitting and then predict the new dose rate given the newly derived calibration for η_i (in keV) with $i := \{30, 40, 50, \dots, 530\}$. The upper integration energy was set to 2800 keV analogue to the previous calculations.

365 We assume that the laboratory-derived \dot{D}_γ are the benchmark value we want reproduce. Figure 11 offers insight in the evolution of \dot{D}_γ for GR1 (blue) and RayMon GR1 (red). The solid line is central value reference for WH2024 from laboratory measurements, and the dashed lines show the 2σ uncertainties (calculation after the Central Dose Model; Galbraith and Roberts (2012)). Setting aside of the calibration-caused discrepancy between the two detectors, both detectors perform similarly between a threshold of ca 30 keV and 200 keV for the counts integration technique and 30 keV to 150 keV for the energy
 370 integration technique and overlap before falling systematically below the laboratory-derived reference value. The uncertainties of all analyses overlap until a threshold of ca 300 keV and it appears that for our setting and tested environment, the output is relatively insensitive to the exact value picked for η (given the uncertainties). The picked values, however, seem to nearly ideally reproduced the benchmark value, while lower values for η as indicated in our experiments would likely overestimate the true \dot{D}_γ .




Figure 11. Estimations for \dot{D}_γ for different values of η for the two investigated detectors. The solid line shows the \dot{D}_γ of the measured site Weiße Hohl derived from laboratory radionuclide estimations and the dashed lines its uncertainties. All uncertainties are shown as 2σ values. For further details see maintext.

375 4 Discussion

Our study reported the performance and dose-rate calibration procedure of two portable semiconductor-based portable γ -ray spectrometers. Both devices host a similar CZT detector that can be operated at ambient temperature, i.e., in situations typical for environmental dose rate measurements as part of trapped-charge dating studies. Unlike literature reporting on γ -ray measurements in the field that used NaI(Tl) or LaBr₃ probes with inch-size diameters, our detectors are considerably smaller (crystal volume 1 cm³), and the systems have a low power consumption, boosting their appeal for trapped-charge dating studies despite that no previous experience was available addressing our field of application. This seems surprising, given the body of available literature about CZT detectors. However, usually, those studies aim at nuclear radiation monitoring (e.g., Alam et al., 2021) or identifying artificial radio-nuclides in environmental studies (e.g., Rahman et al., 2013) for which such detectors are primarily designed.

385 On the plus side, this feature of the detectors simplified the energy/channel calibration with artificial radionuclides because of the detectors' sensitivity to those nuclides. Our energy calibration exhibited peak positions in excellent agreement for both detectors and we concluded that we could apply one single energy calibration. This approach was valid for us, but other detectors likely require separate channel/energy calibration. Although we did not observe a shift in the channel/energy calibration with temperature during all experiments, we highly recommend an energy/channel calibration as part of the post-
390 processing because all subsequent analyses depend on it.

Dating studies require an accurate reading of a sediment matrix's natural γ -radiation field of unknown radionuclide composition in a 4π geometry at the sampling position. Our study proved that both detectors can achieve this in a reasonable time



of 20 min. This value likely works for many environments typically encountered in trapped-charge dating applications. Still, it might be too short for accurate dose rate estimations in settings with a low amount of natural radionuclides or if higher
395 precision is desired. Hence, in case of doubt, measurement times should be adjusted. We recommend a minimum measurement duration of 60 min to obtain the dose-rate calibration curve with a good counting statistics.

A crucial part of our contribution was the determination of the energy threshold η above which the count/energy rate is proportional to the dose rate for natural *radioactive* elements. Given the highly comparable performance characteristics of both detectors, our results can be easily used by others with the same type of detector without repeating all experiments. We tested
400 three different methods (simulation, classical measurements, dose-rate response curve fitting) to determine this threshold and opted for the results from the simulation since the measured natural sites were likely not diverse enough. The cross-check to the natural loess site indicates that the assumption made to simplify the simulation had no significant impact on the results. The threshold found here is considerably lower than results obtained in studies with NaI(Tl) or LaBr₃ probes that place η at ca 300 keV or higher (Mercier and Falguères, 2007; Guérin and Mercier, 2011; Duval and Arnold, 2013). This balances to
405 some extent the lower absolute efficiency of the tiny CZT detectors because it allows exploiting a larger portion of the recorded spectrum.

Nonetheless, it should be kept in mind that all three approaches, simulation, classical experiments, and dose-rate curve fitting, have different meanings. Under the assumption of correct input parameters, the simulation investigates the interaction of the γ -photons with matter for different scenarios and can hence truly determine a range above which the threshold assumption is
410 valid. In other words, the simulation results have merit and provide a solid basis for setting of the thresholds. On the contrary, the experiment findings depend on the matrix composition of the host rock, which in our case, is very similar if translated into relative γ -dose rate contributions from the different radionuclides. We did not observe matching pattern for the threshold from the measurements and the simulation and without simulation, a meaningful determination of η still requires measurements of emitters with pure radionuclide composition, such as the Oxford blocks (Rhodes and Schwenninger, 2007).

415 The threshold quantified in our study is likely not much different for detectors of similar size and with a comparable CZT detector. Therefore we argue that the threshold settings can be adapted if a simulation or a measurement is not possible. This suggestion is further supported by our tests of the Weiße-Hohl measurement with shifting thresholds and minor difference in the detection characteristics will not bias the outcome for the γ -dose rate. Future work should investigate the calibration curve at very low dose rates and in very different environmental settings, since this was not tested in our study.

420 The \dot{D}_γ results of the Weiße-Hohl reflect mainly statistical variations of the different analytical methods. Striking but not puzzling is relatively large coefficient of variation (c_v) of portable CZT detector results (GR1: 11.9%; RayMon GR1: 6.8%) compared to the laboratory measurements (ca 1.7% or lower). Duval and Arnold (2013) reported c_v s around 5% comparable to our laboratory measurements and results own LaBr₃ measurements calibrated at the Clermont-Ferrand sites with more data points typically yielded a c_v s of 5% or better (e.g., Kreutzer et al., 2018a, their Table S9) (typical values: live time: ca 600 s;
425 integrated counts: ca 24000 counts; count rate: 45 s⁻¹)

Furthermore, the larger uncertainties of GR1 compared to RayMon GR1, seem to diminish the overall good performance. For GR1, the weaker performance (larger uncertainties) results from only having three calibration points available, which would



disappear with an additional point. More generally, we argue that this precision can be significantly improved with more points for the dose-rate calibration curve. Those points can be added at any time later, for instance, by measuring more sites around
430 Clermont-Ferrand (Miallier et al., 2009). In such case, however, a check on the energy calibration is mandatory before and after going to the field to monitor potential shifts of the energy spectrum that might be caused by temperature or other technical reasons. Although our results did not encounter such shift, all experiments were carried out in a very short time window.

Finally, what we did not expect of these CZT detectors but should be mention for completeness is that our findings show that those detectors are unsuited for applying the “window” method in environments and for measurements durations typical for our
435 field of application. This would require hour-long measurements to achieve acceptable error margins. For the determination of radionuclide composition laboratory based analytical techniques are unmatched in their effectiveness and precision and they also allow to derive α and β -dose rate components.

5 Conclusions

The primary aim of our study was to test and evaluate the performance of two portable CZT detectors for deployment as active
440 in situ detectors in trapped-charge dating applications. To that end, we measured spectra on natural reference sites with known radionuclide composition in France to derive a dose-rate calibration curves for our two detectors. Background measurements in a low-radiation setting exhibited negligible count rates that can be ignored.

To determine the optimal energy threshold above which the matrix composition of the measured site does not bias the integrated signal to γ -dose rate relation, we performed energy-matter interaction simulations using *GEANT4*. The simulation
445 indicated a suitable energy threshold between 192.5 keV and 242.5 keV for the counts integration technique and 97.5 keV and 222.5 keV for the energy integration technique. We compared those with thresholds derived from cumulative spectra and signal-dose rate regression lines for the two different integration techniques and we found a value of 91 keV for the counting threshold integration and 71 keV for energy counting integration technique. However, given the results from the reference loess site with know radionuclide composition, we discarded the experimentally derived energy thresholds as they are likely too low because
450 of the high similarity of the investigated natural sites. To record a γ -dose rate in typical natural sediment environments, we recommend a measurement time of at least 20 min (this approximates to a total of 4500 counts or better).

A check of our results through measurements at the homogeneous loess deposit near Heidelberg for which we derived the radionuclide composition in the laboratory, confirmed an excellent match of field and laboratory methods, however, with considerably larger (but perhaps more realistic) uncertainties for results from the CZT detectors. Finally, we argue that refined
455 calibrations can further reduce those uncertainties on more sites. Future work may want to extend our calibration curves and explore the performance of the detector in more extreme (low and high) natural radiation fields.

Code and data availability. Raw and partly processed data and R code used in this study is available on Zenodo (<https://doi.org/10.5281/zenodo.13731839>; last access: 2024-09-10).



460 *Author contributions.* SK: Writing – original draft, Validation, Methodology, Formal analysis, Conceptualization, Funding acquisition, Data curation, Software. LM: Writing – review and editing, Methodology, Formal analysis. DM: Writing – review and editing, Resources. NM: Writing – review and editing, Methodology, Formal analysis, Conceptualization.

Competing interests. We declare no conflict of interest and did not receive funds other than what was stated. The Heidelberg Luminescence Laboratory procured the tested systems with its own funds. We contacted the equipment manufacturer to obtain technical information required for the dose-rate simulation and regarding the temperature stability. Other than that, we did not liaise with the manufacturer.

465 *Disclaimer.* Our contribution aims to provide information on the practical use of systems available commercially off-shelf to the trapped-charge dating community and report calibration and performance results. However, we do not favour a particular supplier, and our manuscript must not be understood as a purchase recommendation.

470 *Acknowledgements.* We thank Nicolas Frerebeau for his very responsive support as maintainer of the R package ‘gamma’. We are grateful to Maryam Heydari for performing the measurements at the Weiße Hohl. We further thank Michael Faske from the Scientific Workshop Service of Heidelberg University for designing and printing the strain relief adapter for the GR1 detector and agreeing to share it under CC BY-NC licence conditions. Dennis Gross and Jochen Schreiner enabled the access to FLOSSI and organised the removal of the tree in front of the block.



References

- Adamiec, G. and Aitken, M. J.: Dose-rate conversion factors: update, *Ancient TL*, 16, 37–50, http://ancienttl.org/ATL_16-2_1998/ATL_16-2_Adamiec_p37-50.pdf (last access: 2024-09-02), 1998.
- Agostinelli, S., Allison, J., Amako, K., Apostolakis, J., Araujo, H., Arce, P., Asai, M., Axen, D., Banerjee, S., Barrand, G., Behner, F., Bellagamba, L., Boudreau, J., Broglia, L., Brunengo, A., Burkhardt, H., Chauvie, S., Chuma, J., Chytracek, R., Cooperman, G., Cosmo, G., Degtyarenko, P., Dell'Acqua, A., Depaola, G., Dietrich, D., Enami, R., Feliciello, A., Ferguson, C., Fesefeldt, H., Folger, G., Foppiano, F., Forti, A., Garelli, S., Giani, S., Giannitrapani, R., Gibin, D., Cadenas, J. J. G., González, I., Abril, G. G., Greeniaus, G., Greiner, W., Grichine, V., Grossheim, A., Guatelli, S., Gumplinger, P., Hamatsu, R., Hashimoto, K., Hasui, H., Heikkinen, A., Howard, A., Ivanchenko, V., Johnson, A., Jones, F. W., Kallenbach, J., Kanaya, N., Kawabata, M., Kawabata, Y., Kawaguti, M., Kelner, S., Kent, P., Kimura, A., Kodama, T., Kokoulin, R., Kossov, M., Kurashige, H., Lamanna, E., Lampén, T., Lara, V., Lefebvre, V., Lei, F., Liendl, M., Lockman, W., Longo, F., Magni, S., Maire, M., Medernach, E., Minamimoto, K., de Freitas, P. M., Morita, Y., Murakami, K., Nagamatu, M., Nartallo, R., Nieminen, P., Nishimura, T., Ohtsubo, K., Okamura, M., O'Neale, S., Oohata, Y., Paech, K., Perl, J., Pfeiffer, A., Pia, M. G., Ranjard, F., Rybin, A., Sadilov, S., Salvo, E. D., Santin, G., Sasaki, T., Savvas, N., Sawada, Y., Scherer, S., Sei, S., Sirotenko, V., Smith, D., Starkov, N., Stoecker, H., Sulkimo, J., Takahata, M., Tanaka, S., Tcherniaev, E., Tehrani, E. S., Tropeano, M., Truscott, P., Uno, H., Urban, L., Urban, P., Verderi, M., Walkden, A., Wander, W., Weber, H., Wellisch, J. P., Wenaus, T., Williams, D. C., Wright, D., Yamada, T., Yoshida, H., and Zschesche, D.: GEANT4—a simulation toolkit, *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 506, 250–303, [https://doi.org/10.1016/s0168-9002\(03\)01368-8](https://doi.org/10.1016/s0168-9002(03)01368-8), 2003.
- Aitken, M. J.: *Thermoluminescence dating*, *Studies in archaeological science.*, Academic Press, London, ISBN 978-0-12-046381-7, 1985.
- Alam, M. D., Nasim, S. S., and Hasan, S.: Recent progress in CdZnTe based room temperature detectors for nuclear radiation monitoring, *Progress in Nuclear Energy*, 140, 103918, <https://doi.org/10.1016/j.pnucene.2021.103918>, 2021.
- Alexiev, D., Mo, L., Prokopovich, D. A., Smith, M. L., and Matuchova, M.: Comparison of LaBr₃:Ce and LaCl₃:Ce With NaI(Tl) and Cadmium Zinc Telluride (CZT) Detectors, *IEEE Transactions on Nuclear Science*, 55, 1174–1177, <https://doi.org/10.1109/TNS.2008.922837>, 2008.
- Allaire, J., Xie, Y., Dervieux, C., R Foundation, Wickham, H., Journal of Statistical Software, Vaidyanathan, R., Association for Computing Machinery, Boettiger, C., Elsevier, Broman, K., Mueller, K., Quast, B., Pruim, R., Marwick, B., Wickham, C., Keyes, O., Yu, M., Emaasit, D., Onkelinx, T., Gasparini, A., Desautels, M.-A., Leutnant, D., MDPI, Taylor and Francis, Ögreden, O., Hance, D., Nüst, D., Uvesten, P., Campitelli, E., Muschelli, J., Hayes, A., Kamvar, Z. N., Ross, N., Cannoodt, R., Luguern, D., Kaplan, D. M., Kreuzer, S., Wang, S., Hesselberth, J., and Hyndman, R.: *rticles: Article Formats for R Markdown*, <https://doi.org/10.32614/CRAN.package.rarticles>, r package version 0.27.3, 2024.
- Antoine, P., Rousseau, D.-D., Zöller, L., Lang, A., Munaut, A.-V., Hatté, C., and Fontugne, M.: High-resolution record of the last Interglacial-glacial cycle in the Nussloch loess-palaeosol sequences, Upper Rhine Area, Germany, *Quaternary International*, 76/77, 211–229, 2001.
- Arnold, L. J., Duval, M., Falguères, C., Bahain, J. J., and Demuro, M.: Portable gamma spectrometry with cerium-doped lanthanum bromide scintillators: Suitability assessments for luminescence and electron spin resonance dating applications, *Radiation Measurements*, 47, 6–18, <https://doi.org/10.1016/j.radmeas.2011.09.001>, 2012.
- Baró, J., Sempau, J., Fernández-Varea, J., and Salvat, F.: PENELOPE: An algorithm for Monte Carlo simulation of the penetration and energy loss of electrons and positrons in matter, *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 100, 31–46, [https://doi.org/10.1016/0168-583X\(95\)00349-5](https://doi.org/10.1016/0168-583X(95)00349-5), 1995.



- 510 Bosq, M., Kreutzer, S., Bertran, P., Lanos, P., Dufresne, P., and Schmidt, C.: Last Glacial loess in Europe: luminescence database and chronology of deposition, *Earth System Science Data*, <https://doi.org/10.5194/essd-2023-105>, 2023.
- Bu, M., Murray, A. S., Kook, M., Buylaert, J.-P., and Thomsen, K. J.: The Application of Full Spectrum Analysis to NaI(Tl) Gamma Spectrometry for the Determination of Burial Dose Rates, *Geochronometria*, 48, 161–170, <https://doi.org/10.2478/geochr-2020-0009>, 2021.
- 515 Cresswell, A. J., Carter, J., and Sanderson, D. C. W.: Dose rate conversion parameters: Assessment of nuclear data, *Radiation Measurements*, 120, 195–201, <https://doi.org/10.1016/j.radmeas.2018.02.007>, 2018.
- Duval, M. and Arnold, L.: Field gamma dose-rate assessment in natural sedimentary contexts using LaBr₃(Ce) and NaI(Tl) probes: A comparison between the “threshold” and “windows” techniques, *Applied Radiation and Isotopes*, 74, 36–45, <https://doi.org/10.1016/j.apradiso.2012.12.006>, 2013.
- 520 Frerebeau, N.: khroma: Colour Schemes for Scientific Data Visualization, Université Bordeaux Montaigne, Pessac, France, <https://doi.org/10.5281/zenodo.1472077>, R package version 1.14.0, 2024.
- Frerebeau, N., Lebrun, B., Paradol, G., and Kreutzer, S.: gamma: Dose Rate Estimation from in-Situ Gamma-Ray Spectrometry, Université Bordeaux Montaigne, Pessac, France, <https://doi.org/10.32614/CRAN.package.gamma>, r package version 1.1.0, 2024.
- Galbraith, R. F. and Roberts, R. G.: Statistical aspects of equivalent dose and error calculation and display in OSL dating: An overview and some recommendations, *Quaternary Geochronology*, 11, 1–27, <https://doi.org/10.1016/j.quageo.2012.04.020>, 2012.
- 525 Gilmore, G. R.: Practical gamma-ray spectrometry, John Wiley & sons, Ltd, John Wiley & Sons, Ltd, 2nd edition edn., 2008.
- Godfrey-Smith, D. I., Scallion, P., and Clarke, M. L.: BETA DOSIMETRY OF POTASSIUM FELDSPARS IN SEDIMENT EXTRACTS USING IMAGING MICROPROBE ANALYSIS AND BETA COUNTING, *Geochronometria*, 24, 7–12, http://www.geochronometria.pl/pdf/geo_24/Geo24_2.pdf (last access: 2024-09-02), 2005.
- 530 Guérin, G. and Mercier, N.: Determining gamma dose rates by field gamma spectroscopy in sedimentary media: Results of Monte Carlo simulations, *Radiation Measurements*, 46, 190–195, <https://doi.org/10.1016/j.radmeas.2010.10.003>, 2011.
- Guérin, G., Mercier, N., and Adamiec, G.: Dose-rate conversion factors: update, *Ancient TL*, 29, 5–8, http://ancienttl.org/ATL_29-1_2011/ATL_29-1_Guerin_p5-8.pdf (last access: 2024-09-02), 2011.
- Guérin, G., Mercier, N., Nathan, R., Adamiec, G., and Lefrais, Y.: On the use of the infinite matrix assumption and associated concepts: A critical review, *Radiation Measurements*, 47, 778–785, <https://doi.org/10.1016/j.radmeas.2012.04.004>, 2012.
- 535 Guibert, P. and Schvoerer, M.: TL dating: Low background gamma spectrometry as a tool for the determination of the annual dose, *International Journal of Radiation Applications and Instrumentation. Part D. Nuclear Tracks and Radiation Measurements*, 18, 231–238, [https://doi.org/10.1016/1359-0189\(91\)90117-z](https://doi.org/10.1016/1359-0189(91)90117-z), 1991.
- Guérin, G. and Mercier, N.: Preliminary insight into dose deposition processes in sedimentary media on a scale of single grains: Monte Carlo modelling of the effect of water on the gamma dose rate, *Radiation Measurements*, 47, 541 – 547, <https://doi.org/10.1016/j.radmeas.2012.05.004>, 2012.
- 540 Hossain, S. M., Corte, F. D., Vandenberghe, D., and den haute, P. V.: A comparison of methods for the annual radiation dose determination in the luminescence dating of loess sediment, *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 490, 598–613, [https://doi.org/10.1016/s0168-9002\(02\)01078-1](https://doi.org/10.1016/s0168-9002(02)01078-1), 2002.
- 545 Hutton, J. T. and Prescott, J. R.: Field and laboratory measurements of low-level thorium, uranium and potassium, *International Journal of Radiation Applications and Instrumentation. Part D. Nuclear Tracks and Radiation Measurements*, 20, 367–370, [https://doi.org/10.1016/1359-0189\(92\)90066-5](https://doi.org/10.1016/1359-0189(92)90066-5), 1992.



- Ivanchenko, V., Apostolakis, J., Bagulya, A., Abdelouahed, H. B., Black, R., Bogdanov, A., Burkhard, H., Chauvie, S., Cirrone, P., Cuttone, G., Depaola, G., Rosa, F. D., Elles, S., Francis, Z., Grichine, V., Gumplinger, P., Gueye, P., Incerti, S., Ivanchenko, A., Jacquemier, J.,
550 Lechner, A., Longo, F., Kadri, O., Karakatsanis, N., Karamitros, M., Kokoulin, R., Kurashige, H., Maire, M., Mantero, A., Mascialino, B., Moscicki, J., Pandola, L., Perl, J., Petrovic, I., Ristic-Fira, A., Romano, F., Russo, G., Santin, G., Schaelicke, A., Toshito, T., Tran, H., Urban, L., Yamashita, T., and Zacharatou, C.: Recent Improvements in Geant4 Electromagnetic Physics Models and Interfaces, *Progress in Nuclear Science and Technology*, 2, 898–903, <https://doi.org/10.15669/pnst.2.898>, 2011.
- Kalchgruber, R.: α -Al₂O₃:C als dosimeter zur bestimmung der dosisleistung bei der lumineszenzdatierung, Ph.D. thesis, Heidelberg University, 2002.
555
- Kalchgruber, R. and Wagner, G. A.: Separate assessment of natural beta and gamma dose-rates with TL from single-crystal chips, *Radiation Measurements*, 41, 154–162, <https://doi.org/10.1016/j.radmeas.2005.04.002>, 2006.
- Kalchgruber, R., Fuchs, M., Murray, A. S., and Wagner, G. A.: Evaluating dose-rate distributions in natural sediments using α -Al₂O₃:C grains, *Radiation Measurements*, 37, 293–297, [https://doi.org/10.1016/S1350-4487\(03\)00012-X](https://doi.org/10.1016/S1350-4487(03)00012-X), 2003.
- 560 Kolb, T., Tudyka, K., Kadereit, A., Lomax, J., Poreba, G., Zander, A., Zipf, L., and Fuchs, M.: The μ Dose system: determination of environmental dose rates by combined alpha and beta counting – performance tests and practical experiences, *Geochronology*, 4, 1–31, <https://doi.org/10.5194/gchron-4-1-2022>, 2022.
- Kreutzer, S., Duval, M., Bartz, M., Bertran, P., Bosq, M., Eynaud, F., Verdin, F., and Mercier, N.: Deciphering long-term coastal dynamics using IR-RF and ESR dating: A case study from Médoc, south-West France, *Quaternary Geochronology*, 48, 108–120,
565 <https://doi.org/10.1016/j.quageo.2018.09.005>, 2018a.
- Kreutzer, S., Martin, L., Guérin, G., Tribolo, C., Selva, P., and Mercier, N.: Environmental dose rate determination using a passive dosimeter: Techniques and workflow for alpha-Al₂O₃:C chips, *Geochronometria*, 45, 56–67, <https://doi.org/10.1515/geochr-2015-0086>, 2018b.
- Kreutzer, S., Martin, L., Miallier, D., Mercier, N., and Faske, M.: Dataset: Environmental Gamma Dose Rate Measurements using CZT Detectors, <https://doi.org/10.5281/zenodo.13731839>, 2024.
- 570 Lebrun, B., Frerebeau, N., Paradol, G., Guérin, G., Mercier, N., Tribolo, C., Lahaye, C., and Magali, R.: gamma: An R package for dose rate estimation from in-situ gamma-ray spectrometry measurements, *Ancient TL*, 38, 11–15, http://ancienttl.org/ATL_38-2_2020/ATL_38-2_Lebrun_p1-5.pdf (last access: 2024-09-02), 2020.
- Limousin, O.: New trends in CdTe and CdZnTe detectors for X- and gamma-ray applications, *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 504, 24–37, [https://doi.org/10.1016/S0168-9002\(03\)00745-9](https://doi.org/10.1016/S0168-9002(03)00745-9), 2003.
575
- Liritzis, I., Stamoulis, K., Papachristodoulou, C., and Ioannides, K.: A re-evaluation of radiation dose-rate conversion factors, *Mediterranean Archaeology and Archaeometry*, 12, 1–15, <http://maajournal.com/Issues/2012/pdf/FullTextLiritzis.pdf> (last access: 2024-09-02), 2013.
- Løvborg, L. and Kirkegaard, P.: Response of 3'' × 3'' NaI(Tl) detectors to terrestrial gamma radiation, *Nuclear Instruments and Methods*, 121, 239–251, [https://doi.org/10.1016/0029-554x\(74\)90072-x](https://doi.org/10.1016/0029-554x(74)90072-x), 1974.
- 580 Martin, L., Incerti, S., and Mercier, N.: DosiVox: Implementing Geant 4-based software for dosimetry simulations relevant to luminescence and ESR dating techniques, *Ancient TL*, 33, 1–10, http://ancienttl.org/ATL_33-1_2015/ATL_33-1_Martin_p1-10.pdf (last access: 2024-10-24), 2015.
- Martin, L., Duval, M., and Arnold, L. J.: To what extent do field conditions affect gamma dose rate determination using portable gamma spectrometry?, *Radiation Physics and Chemistry*, 216, 111 365, <https://doi.org/10.1016/j.radphyschem.2023.111365>, 2024.



- 585 Mauz, B., Hubmer, A., Bahl, C., Lettner, H., and Lang, A.: Comparing two efficiency calibration methods used in gamma spectrometry, *Ancient TL*, 39, 12–17, http://ancienttl.org/ATL_39-2_2021/ATL_39-2_Mauz_p12-17.pdf (last access: 2024-09-02), 2021.
- Mercier, N. and Falguères, C.: Field gamma dose-rate measurement with a NaI(Tl) detector: re-evaluation of the technique, *Ancient TL*, 25, 1–4, http://ancienttl.org/ATL_25-1_2007/ATL_25-1_Mercier_p1-4.pdf (last access: 2024-09-02), 2007.
- Miallier, D., Guérin, G., Mercier, N., Pilleyre, T., and Sanzelle, S.: The Clermont radiometric reference rocks: a convenient tool for dosimetric purposes, *Ancient TL*, 27, 37–44, http://ancienttl.org/ATL_27-2_2009/ATL_27-2_Miallier_p37-44.pdf (last access: 2024-09-02), 2009.
- 590 Murray, A., Bownman, S., and Aitken, M.: Evaluation of the gamma dose-rate contribution, *PACT*, 2, 84–96, 1978.
- Preusser, F. and Kasper, H. U.: Comparison of dose rate determination using high-resolution gamma spectrometry and inductively coupled plasma - mass spectrometry, *Ancient TL*, 19, 19–23, http://ancienttl.org/ATL_19-1_2001/ATL_19-1_Preusser_p19-23.pdf (last access: 2024-09-02), 2001.
- 595 R Core Team: R: A language and environment for statistical computing, <https://r-project.org/> (last access: 2024-09-02), 2024.
- Rahman, R., Plater, A. J., Nolan, P. J., and Appleby, P. G.: Assessing CZT detector performance for environmental radioactivity investigations, *Radiation Protection Dosimetry*, 154, 477–482, <https://doi.org/10.1093/rpd/ncs253>, 2013.
- Rhodes, E. J. and Schwenninger, J. L.: Dose rates and radioisotope concentrations in the concrete calibration blocks at Oxford, *Ancient TL*, 25, 5–8, http://ancienttl.org/ATL_25-1_2007/ATL_25-1_Rhodes_p5-8.pdf (last access: 2024-09-02), 2007.
- 600 Richter, D., Dombrowski, H., Neumaier, S., Guibert, P., and Zink, A. C.: Environmental gamma dosimetry with OSL of -Al₂O₃:C for in situ sediment measurements, *Radiation Protection Dosimetry*, 141, 27–35, <https://doi.org/10.1093/rpd/ncq146>, 2010.
- Rieser, U.: Low-level gamma-spektrometrie zum zwecke der dosisleistungsbestimmung bei der lumineszenz datierung, Diplomarbeit (unpublished), Ruprecht-Karls University of Heidelberg, Heidelberg, 1991.
- Sanderson, D. C. W. and Murphy, S.: Using simple portable OSL measurements and laboratory characterisation to help understand complex and heterogeneous sediment sequences for luminescence dating, *Quaternary Geochronology*, 5, 299–305, <https://doi.org/10.1016/j.quageo.2009.02.001>, 2010.
- 605 Scheiber, C. and Chambron, J.: CdTe detectors in medicine: a review of current applications and future perspectives, *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 322, 604–614, [https://doi.org/10.1016/0168-9002\(92\)91239-6](https://doi.org/10.1016/0168-9002(92)91239-6), 1992.
- 610 Tudyka, K., Miłosz, S., Adamiec, G., Bluszcz, A., Poręba, G., Paszkowski, Ł., and Kolarczyk, A.: μ Dose_ A compact system for environmental radioactivity and dose rate measurement, *Radiation Measurements*, 118, 8–13, <https://doi.org/10.1016/j.radmeas.2018.07.016>, 2018.
- Tudyka, K., Kłosok, K., Gosek, M., Kolarczyk, A., Miłosz, S., Szymak, A., Pilśniak, A., Moska, P., and Poręba, G.: μ DOSE+: Environmental radioactivity and dose rate measurement system with active shielding boosted by machine learning, *Measurement*, 234, 114 854, <https://doi.org/10.1016/j.measurement.2024.114854>, 2024.
- 615 Verger, L., Bonnefoy, J. P., Glasser, F., and Ouvrier-Buffet, P.: New Developments in CdTe and CdZnTe Detectors for X and y-Ray Applications, *Journal of Electronic Materials*, 26, 738–744, <https://doi.org/10.1007/s11664-997-0225-2>, 1997.
- Vermeesch, P.: IsoplotR: a free and open toolbox for geochronology, *Geoscience Frontiers*, 9, 1479–1493, <https://doi.org/10.1016/j.gsf.2018.04.001>, 2018.
- 620 Wickham, H.: ggplot2: Elegant Graphics for Data Analysis, Springer-Verlag New York, ISBN 978-3-319-24277-4, <https://ggplot2.tidyverse.org> (last access: 2024-09-02), 2016.



York, D., Evensen, N. M., Martínez, M. L., and De Basabe Delgado, J.: Unified equations for the slope, intercept, and standard errors of the best straight line, *American Journal of Physics*, 72, 367–375, <https://doi.org/10.1119/1.1632486>, 2004.

Zöller, L. and Pernicka, E.: A note on overcounting in alpha-counters and its elimination, *Ancient TL*, 7, 11–14, http://ancienttl.org/ATL_07-1_1989/ATL_07-1_Zoller_p11-14.pdf (last access: 2024-09-02), 1989.

625